

## New England Biolabs Certificate of Analysis

**Product Name:** XbaI  
**Catalog Number:** R0145M  
**Concentration:** 100,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (dam-/HindIII digest) in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10107526  
**Expiration Date:** 05/2023  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0145T/M v1.0

XbaI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0145M VIAL	XbaI	10107526	Pass
B7024A VIAL	Gel Loading Dye, Purple (6X)	10093123	Pass
B6004S VIAL	rCutSmart™ Buffer	10107577	Pass

Assay Name/Specification	Lot # 10107526
<b>Protein Purity Assay (SDS-PAGE)</b> XbaI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda HindIII dam- DNA and a minimum of 200 Units of XbaI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of Adenovirus-2 DNA with XbaI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with XbaI.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 200 units of XbaI incubated for 4	Pass


Assay Name/Specification	Lot # 10107526
hours at 37°C releases <0.1% of the total radioactivity.	
<p><b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 100 Units of XbaI incubated for 4 hours at 37°C results in &lt;10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	<b>Pass</b>
<p><b>Blue-White Screening (Terminal Integrity)</b> A sample of pUC19 vector linearized with a 10-fold excess of XbaI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in &lt;1% white colonies.</p>	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Penghua Zhang  
Production Scientist  
08 Jun 2021



Michael Tonello  
Packaging Quality Control Inspector  
08 Jun 2021