

New England Biolabs Certificate of Analysis

Product Name: *Mlul*
Catalog Number: *R0198L*
Concentration: *10,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot Number: *10023147*
Expiration Date: *09/2020*
Storage Temperature: *-20°C*
Storage Conditions: *100 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0198S/L v1.0*

Mlul Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0198LVIAL	Mlul	10023148	Pass
B7203SVIAL	NEBuffer™ 3.1	10010189	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10018419	Pass

Assay Name/Specification	Lot # 10023147
<p>Endonuclease Activity (Nicking) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled pUC19 DNA and a minimum of 30 units of Mlul incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of Mlul incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Mlul, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Mlul.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 100 Units of Mlul incubated for 16 hours at 37°C results in a DNA pattern free of</p>	Pass

Assay Name/Specification	Lot # 10023147
detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
06 Jun 2018



Michael Tonello
Packaging Quality Control Inspector
01 Oct 2018