

New England Biolabs Certificate of Analysis

Product Name: *FauI*
Catalog #: *R0651S/L*
Concentration: *5,000 units/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 55°C in a total reaction volume of 50 µl.*
Lot #: *0021707*
Assay Date: *07/2017*
Expiration Date: *7/2019*
Storage Temp: *-20°C*
Storage Conditions: *50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0651S/L v1.0*
Effective Date: *11 Jun 2013*

Assay Name/Specification (minimum release criteria)	Lot #0021707
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 5 units of FauI incubated for 4 hours at 55°C releases <1.0% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 5-fold over-digestion of Lambda DNA with FauI, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with FauI.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 5 Units of FauI incubated for 16 hours at 55°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by
Derek Robinson
11 Jun 2013



Inspected by
Jianying Luo
18 Jul 2017

