

## New England Biolabs Certificate of Analysis

**Product Name:** BtsI-v2  
**Catalog Number:** R0667S  
**Concentration:** 10,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10108460  
**Expiration Date:** 03/2023  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM KCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 200 µg/ml BSA (pH 7.4 @ 25°C)  
**Specification Version:** PS-R0667S/L v3.0

BtsI-v2 Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0667SVIAL	BtsI-v2	10100838	Pass
B6004SVIAL	rCutSmart™ Buffer	10108412	Pass

Assay Name/Specification	Lot # 10108460
<b>Protein Purity Assay (SDS-PAGE)</b> BtsI-v2 is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of Lambda DNA with BtsI-v2, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with BtsI-v2.	Pass
<b>Functional Testing (15 minute Digest)</b> A 50 µl reaction in CutSmart® Buffer containing 1 µg of Lambda DNA and 1 µl of BtsI-v2 incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart® Buffer containing 1 µg of Lambda DNA and a minimum of 10 units of BtsI-v2 incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Assay Name/Specification	Lot # 10108460
<p><b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart® Buffer containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 50 units of BtsI-v2 incubated for 4 hours at 37°C releases &lt;0.1% of the total radioactivity.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.




---

Penghua Zhang  
Production Scientist  
17 Jun 2021




---

Jason Davis  
Packaging Quality Control Inspector  
17 Jun 2021