

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Nsil-HF®
Catalog Number: R3127L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Lot Number:10032008Expiration Date:12/2020Storage Temperature:-20°C

Storage Conditions: 300 mM NaCl , 10 mM Tris-HCl (pH 7.4), 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol , 500 μg/ml BSA

Specification Version: PS-R3127S/L v1.0

Nsil-HF® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R3127LVIAL	Nsil-HF®	10032009	Pass	
B7204SVIAL	CutSmart® Buffer	10021125	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10021135	Pass	

Assay Name/Specification	Lot # 10032008
Endonuclease Activity (Nicking) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of supercoiled φX174 DNA and a	Pass
minimum of 100 units of Nsil-HF incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of Nsil-HF incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Test (15 minute Digest) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and 1 µl of Nsil-HF incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with Nsil-HF, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments,	Pass



R3127L / Lot: 10032008 Page 1 of 2 This product has been tested and shown to be in compliance with all specifications.

Tony Spear-Alfonso Production Scientist

10 Dec 2018

Michael Tonello

Packaging Quality Control Inspector

20 Dec 2018

