

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Nsil-HF®
Catalog Number: R3127S
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Lot Number: 10046681
Expiration Date: 06/2021
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl , 10 mM Tris-HCl (pH 7.4), 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol , 500 μg/ml BSA

Specification Version: PS-R3127S/L v1.0

Nsil-HF® Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R3127SVIAL	Nsil-HF®	10046682	Pass	
B7204SVIAL	CutSmart® Buffer	10043915	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10043349	Pass	

Assay Name/Specification	Lot # 10046681
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with Nsil-HF, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Nsil-HF.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 units of Nsil-HF incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) Nsil-HF is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled \$\phi X174 DNA\$ and a minimum of 100 units of Nsil-HF incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass



R3127S / Lot: 10046681

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

David Guo

**Production Scientist** 

20 Jun 2019

Jay Minichiello

Packaging Quality Control Inspector

09 Jul 2019



R3127S / Lot: 10046681

Page 2 of 2