

## New England Biolabs Certificate of Analysis

**Product Name:** PvuII-HF<sup>®</sup>  
**Catalog Number:** R3151S  
**Concentration:** 20,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37° in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10101097  
**Expiration Date:** 03/2023  
**Storage Temperature:** -20°C  
**Storage Conditions:** 200 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R3151S/L v1.0

PvuII-HF <sup>®</sup> Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3151SVIAL	PvuII-HF <sup>®</sup>	10101098	Pass
B7204SVIAL	CutSmart <sup>®</sup> Buffer	10097266	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10091035	Pass

Assay Name/Specification	Lot # 10101097
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart <sup>™</sup> Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of PvuII-HF <sup>™</sup> incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart <sup>™</sup> Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 200 units of PvuII-HF <sup>™</sup> incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in CutSmart <sup>™</sup> Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 200 Units of PvuII-HF <sup>™</sup> incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of Lambda DNA with PvuII-HF <sup>™</sup> , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated	Pass

Assay Name/Specification	Lot # 10101097
fragments, >95% can be recut with PvuII-HF™.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

  
Penghua Zhang  
Production Scientist  
17 Mar 2021

  
Josh Hersey  
Packaging Quality Control Inspector  
17 Mar 2021