


New England Biolabs Certificate of Analysis

Product Name: Monarch RNA Purification Columns
Catalog Number: T2007L
Lot Number: 10043830
Expiration Date: 01/2021
Storage Temperature: 25°C
Specification Version: PS-T2007L v1.0

Assay Name/Specification	Lot # 10043830
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® RNA Purification Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
 Production Scientist
 05 Sep 2019



Heidi Church
 Packaging Quality Control Inspector
 05 Sep 2019