


New England Biolabs Certificate of Analysis

Product Name: *Monarch RNA Cleanup Columns (10 µg)*
 Catalog Number: *T2037L*
 Packaging Lot Number: *10023556*
 Expiration Date: *07/2020*
 Storage Temperature: *25°C*
 Specification Version: *PS-T2037L v1.0*

Assay Name/Specification	Lot # 10023556
<p>Functional Testing (RNA Recovery and Purity) Twenty-four Monarch® RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of rRNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
19 Jun 2019



Heidi Church
Packaging Quality Control Inspector
19 Jun 2019