

## New England Biolabs Certificate of Analysis


**Product Name:** *Monarch<sup>®</sup> RNA Cleanup Columns (10 µg)*  
**Catalog Number:** *T2037L*  
**Packaging Lot Number:** *10180851*  
**Expiration Date:** *03/2024*  
**Storage Temperature:** *25°C*  
**Specification Version:** *PS-T2037L v3.0*

Monarch <sup>®</sup> RNA Cleanup Columns (10 µg) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2037-1	RNA Cleanup 50 Columns (10 µg)	10137053	<b>Pass</b>
T2018-1	Monarch <sup>®</sup> Collection Tubes II	10137027	<b>Pass</b>


Assay Name/Specification	Lot # 10180851
<b>Functional Testing (RNA Recovery and Purity)</b> Twenty-four Monarch <sup>®</sup> RNA Cleanup Columns (T2037, 10 µg) are tested with 1 µg of RNA resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Ana Egana  
Production Scientist  
13 Feb 2023



Nikki Samuelson  
Packaging Quality Control Inspector  
13 Feb 2023